


# ***Declaration of Conformity***

<b>MANUFACTURER:</b>	<b>i-SENS, Inc.</b> <b>43, Banpo-daero 28-gil, Seocho-gu,</b> <b>Seoul 06646, Korea</b>
<b>MANUFACTURING FACILITY:</b>	<b>i-SENS Wonju Factory</b> <b>94-1, Donghwagongdan-ro, Munmak-eup,</b> <b>Wonju-si, Gangwon-do 26365, Korea</b>
<b>EUROPEAN REPRESENTATIVE:</b>	<b>Medical Technology Promedt Consulting GmbH</b> <b>Altenhofstrasse 80, 66386 St. Ingbert, Germany</b>
<b>PRODUCT:</b>	<b>CareSens PRO Blood Glucose Monitoring System</b>
<b>MODEL:</b>	<b>See List of Products</b>
<b>CLASSIFICATION:</b>	<b>List B according to Annex II of IVDD</b>
<b>CONFORMITY ASSESSMENT ROUTE:</b>	<b>IVDD ANNEX IV without section 4 and 6</b> <b>Applied</b>

**We herewith declare that the above-mentioned products meet the provision of the Council Directive 98/79/EC for in vitro diagnostic medical devices under the exclusive responsibility of manufacturer. All supporting documentation is retained at the premises of the manufacturer.**

<b>STANDARD APPLIED:</b>	<b>See List of Applied Standards</b>
<b>NOTIFIED BODY:</b>	<b>TÜV SÜD PRODUCT SERVICE GmbH</b> <b>Ridlerstraße 65, 80339 Munich, Germany</b> <b>(Notified Body Number 0123)</b>
<b>CERTIFICATE:</b>	<b>V1 090700 0032 Rev. 01</b> <b>(Valid until: 2024-04-15)</b>
<b>START OF CE-MARKING:</b>	<b>See List of Products</b>
<b>PLACE, DATE OF ISSUE:</b>	<b>Seoul, 2022-11-01</b>
<b>SIGNATURE:</b>	

---

**CEO**  
**Geun Sig Cha**

## **List of Products**

**Brand Name/Model**

CareSens PRO Blood Glucose Monitoring System, Model GM01HAA

- CareSens PRO Blood Glucose Meter, EDMA: 21 06 01
- CareSens PRO Blood Glucose Test Strips, EDMA: 11 70 01 01 00
- CareSens PRO Glucose Control Solutions, EDMA: 11 50 90 90 00

**\* Start of CE Marking: 2015-08-03**

## **List of Applied Standards**

<b>Document Number</b>	<b>Title of Document</b>
EN ISO 13485: 2016	Medical devices - Quality management systems -Requirements for regulatory purposes
EN ISO 18113-1: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 1: Terms, definitions and general requirements
EN ISO 18113-4: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 4: In vitro diagnostic reagents for self-testing
EN ISO 18113-5: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 5: In vitro diagnostic instruments for self-testing
EN ISO 15223-1: 2016	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements
EN 13532: 2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612: 2002	Performance evaluation of in vitro diagnostic medical devices
EN ISO 23640: 2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents
EN ISO 14971: 2012	Medical devices - Application of risk management to medical devices
EN ISO 15197: 2015	In vitro diagnostic test systems - Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus
EN ISO 17511: 2003	In vitro diagnostic medical devices - Measurements of quantities in samples of biological origin - Metrological traceability of values assigned to calibrators and control materials
EN 61010-1: 2001 Second edition	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements
EN 61010-2-101: 2002	Safety requirements for electrical equipment for measurement, control and laboratory use. Particular requirements for in vitro diagnostic (IVD) medical equipment
EN 61326-1: 2006	Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements
EN 61326-2-6: 2008	Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. In vitro diagnostic (IVD) medical equipment
EN 60068-2-64: 2008	Environmental testing. Tests. Test Fh. Vibration, broadband random and guidance
EN 62304: 2006	Medical device software - Software life cycle processes
EN 62366: 2008	Medical devices - Application of usability engineering to medical devices